



Radiometric uncertainty analysis of the GLAMR calibration facility

GLAMR: Goddard Laser for Absolute Measurement of Radiance

Brendan McAndrew^[a], Julia Barsi^[b], Andrei Sushkov^[c], and Joel McCorkel^[a]

^[a] NASA Goddard Space Flight Center

^[b] Science Systems and Applications, Inc.

^[c] Genesis Engineering, Inc.




GLAMR

Basic calibration scheme

NIST Primary Optical Watt Radiometer (POWER)

electrical substitution optical power meter


Optical watts referenced to electrical watts



NIST reference radiometer



Highly uniform detector, precision apertures

watts \rightarrow watts/(m²*sr)





Basic calibration scheme

Goddard/instrument manufacturer integration and test facility



GLAMR

Transfer radiometer absolute calibration




Calibration is based on the electrical substitution method: optical power is matched to electrical power in a bolometer.

Transfer radiometers and associated amplifiers are sent to NIST periodically for absolute radiance calibration

Radiometers are used in a set of three: silicon, InGaAs, and extended InGaAs to cover the spectral range 310 nm to 2500 nm

Temperature stabilized, unbiased photovoltaic mode of operation



Transfer radiometers

GLAMR

Integrating Sphere Uniformity


Sphere uniformity measured with combination of fixed radiometers and a narrow field of view scanning radiometer

Scanning radiometer mounted on X-Y- θ motion stages: 2 axis linear and 1 axis rotation

Position scanned in grid pattern with 20 mm step size


Center position repeated every other measurement to account for drift (scanning radiometer not temperature stabilized)

Measurements repeated at several illumination wavelengths between 345 nm and 2000 nm



Scanning radiometer


Fixed radiometers
(sphere monitors)



Stage axis directions

Scanning radiometer detail

Radiometer design: interchangeable Si and InGaAs photodiodes with 0.75° field of view Gershun tube



Photodiode with Gershun tube

Non-imaging, no refractive optics


Unbiased, photovoltaic mode operation, same as sphere monitors

Photodiodes



One square (Si) and one round (XIGA) photodiode


Detector	Active area	Size	
Silicon	4.5 mm ²	2.12 × 2.12 mm	Si FOV 6% larger
XIGA	3.14 mm ²	∅ 2mm	




Field of view note

Spatial resolution:
area of aperture 1

Angular resolution:
 $\theta = \tan^{-1}(d/L)$, field of view = 2θ




Field of view for 2 mm diameter apertures




0.75° FWHM ($\pm 0.375^\circ$) Gershun tube length 93 mm

50% area overlap




0% area overlap



GLAMR


Uniformity measurement setup



Fibers

Sphere monitors

Scanning radiometer





Laboratory configuration

GLAMR


Scan sequence

- Automated spatial scan and data collection at constant wavelength and radiance
- 5 Hz radiance data acquisition
- 45 seconds between stage commands
- 26 seconds minimum dwell time at each position




Detail position stepping

Stage position





Center position (0,0) repeated every other measurement






GLAMR

Spatial uniformity data




Variation $\pm 0.1\%$ horizontally,
 $\pm 0.5\%$ vertically

No dependence with wavelength




GLAMR

Combined data sets



Sphere close to symmetric left to right, brightest in the middle




Sphere brightest at the top

X-direction average: each point is an average of measurements taken along the same vertical column

Y-direction average: each point is an average of measurements taken along the same horizontal row

Stage axis directions




GLAMR


Angular uniformity data

Uniform to 0.1% within $\pm 10^\circ$
No dependence with wavelength

Angle measured from port center to sphere relative to port centerline



Sphere Angular Uniformity
angle and x translation



GLAMR


Fluorescence

For the GLAMR sphere, average reflectivity $\bar{\rho} = 0.95$


$$[\rho_{\text{Sphere material}} * (1 - A_{\text{port}}/A_{\text{sphere}})]$$

Average number of reflections $r_{\text{avg}} = \frac{-1}{\ln \bar{\rho}} \approx 20$


Fluorescence signal scales with average number of reflections, in our case a factor of 20



Integrating sphere illuminated with fiber coupled monochromatic light, linewidth < 1 nm




Fluorescence should be observable as light at longer wavelength than incident light and broader linewidth



Fluorescence data acquired with ASD spectro-radiometer





1. Unfiltered data acquired at short exposure time 272 ms
2. Long pass filter data acquired at long exposure time 1.16 minutes
3. Raw data dark subtracted and converted to counts/ms
4. ASD calibration and filter transmission applied
5. Filtered and unfiltered data combined to extend dynamic range
6. Fluorescence *spectral radiance* normalized to laser wavelength *radiance*



ASD fore optic and filter mount in front of integrating sphere

GLAMR


Example data



Applying calibration and normalizing


Fluorescence normalized to radiance of illumination wavelength by integrating area of peak at laser wavelength over ASD bandwidth

Compiled fluorescence data, 4/19 and 5/6 2021



illumination wavelength — 341nm — 360nm — 380nm — 400nm — 420nm


Compiled fluorescence data, 11/28 2022



— 341.2nm — 360nm — 380nm — 400nm — 420nm

Linearity

Transfer radiometer linearity checked in NIST beam conjoiner facility (H. Yoon) using superposition principle



NIST beam conjoiner schematic

Silicon radiometer
850 nm bandpass filter
Broadband <1000 nm


InGaAs and ext. InGaAs radiometers
broadband >1000 nm

No detectable nonlinearity found in transfer radiometers


Radiometer illuminated by light from two different paths with identical spectrum and field of view: flux from combined paths equals sum of flux from each path separately.

Sphere monitor linearity


IGA SM/IGA TR 1500 nm




XIGA SM/XIGA TR 1500 nm




IGA SM/IGA TR 970 nm



IGA SM/IGA TR 1615 nm









XIGA SM/XIGA TR 1615 nm




InGaAs and extended InGaAs sphere monitors linear over all conditions tested, 970 nm – 1615 nm

Sphere monitor linearity

Slight wavelength dependent nonlinearity in silicon sphere monitor




Custom radiometer mount co-aligns transfer radiometer fields of view on integrating sphere



Radiometer mount with alignment guide


Sphere monitors



Calibration setup, alignment guide removed


Transfer radiometers

Comparison of OCI Sphere Calibrations



Transfer radiometer to sphere monitor ratio

- Si OCI_WindowSphereCal_Jan2023
- IGA OCI_SphereCal_Aug2021
- XIGA OCI_WindowSphereCal_Jan2023
- Si OCI_WindowSphereCal_Jun2022_L2
- IGA OCI_SphereCal_Dec2021_L2
- XIGA OCI_WindowSphereCal_Jun2022_L2



Ratio of repeat sphere calibrations

- Si Jun2022:Jan2023 (window)
- IGA Aug2021:Dec2021
- XIGA Jun2022:Jan2023 (window)

Wavelength Range	UV 350nm-400nm	VIS 400nm-820nm	NIR 865nm	NIR 940nm	NIR 1038nm	NIR 1250nm	NIR 1378nm	SWIR 1615nm	SWIR 2130nm	SWIR 2260nm
Combined uncertainty terms (root sum square)	0.23%	0.17%	0.15%	0.15%	0.38%	0.35%	1.06%	0.35%	0.36%	0.34%
Transfer radiometer absolute calibration	0.20%	0.13%	0.10%	0.10%	0.36%	0.33%	0.33%	0.33%	0.34%	0.32%
Sphere repeatability	0.04%	0.04%	0.04%	0.04%	0.04%	0.04%	1.0%	0.04%	0.04%	0.04%
Measurement noise	0.01%	0.01%	0.01%	0.01%	0.01%	0.01%	0.01%	0.01%	0.01%	0.01%
Uniformity of the sphere	0.10%	0.10%	0.10%	0.10%	0.10%	0.10%	0.10%	0.10%	0.10%	0.10%

Instrument calibration considerations

Aperture and field of view

Dynamic range


Frequency bandwidth and sample rate

Measurement averaging time

Out of band fluorescence correction

The logo for the GLAMR (Global Light and Radiometer Measurement) instrument, featuring the acronym in white capital letters on a dark background.

Questions?



Integrating sphere timelapse during a radiometer calibration